

Agence Nord: ZA Object'Ifs Sud - Lot A3 6 Allée Emilie du Châtelet 14123 Ifs tél : 02.31.34.50.74 fax : 02.31.34.55.17



Agence Est: Parc Club des Tanneries 2 Rue de la Faisanderie 67380 Lingolsheim tél : 03.88.04.01.81 fax : 03.68.93.01.52

Agence Sud:

Bât Le Venango. 392 Rue Jean Dausset AGROPARC - BP11575 84916 Avignon Cédex 9 tél : 04.90.27.17.95 fax : 04.90.27.17.52 www.deltalabo.fr info@deltalabo.fr

Herschel Emulsifier 96700-2

Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412



The Compact and Efficient Solution

- 4 independently controlled test stations
- Multiple operator safety features
- Motorised raising and lowering
- Automated test sequence
- Intermediate scraping position
- Integrated timing
- Easy handling and removal of sample
- No removal of the paddles required
- Guaranteed paddle rotation speed
- Non-reflective enhanced LED lighting
- Large viewing window
- Large LCD touchscreen display
- Small footprint
- Intuitive software package

Petroleum Oils • Synthetic Fluids

Herschel Emulsifier 96700-2

Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412

Innovative Design

The Seta Herschel Emulsifier is a compact and efficient benchtop instrument designed to measure the ability of petroleum oils and synthetic fluids to separate from water. The instrument incorporates 4 test stirrers, with independent control and motorised raising and lowering. Samples can be tested simultaneously or individually to suit laboratory requirements.

Each stirrer is held by a self centering collet to ensure concentricity within 1mm and can remain in situ when removing or replacing cylinders. Positioning within 6mm from the bottom of the cylinder is automatically achieved using a motorised actuator.

After the stirring period the motorised head raises the stirrer until clear of the graduate cylinder to allow for wiping and on completion of the test, the stirrer will revert to the higher, home position.

Test cylinders are located in a temperature controlled bath with an adjustable set point of either 54 or 82°C. A large viewing window and non-reflective LED lighting assists operator reading. The bath top plate is designed to allow cylinders to tilt for easy removal. A drain valve is provided for service and maintenance.

A large colour touch screen display is used to initiate tests and provides automated sequencing with an audible and visual reminder at each recording interval. Custom test parameters are also user adjustable.

Operator Interface and Integrated Timing



Typical Applications

• Petroleum Oils, Synthetic Fluids

Test Method

ASTM D1401. A sample is mixed with water, heated in the bath and stirred for 5 minutes whilst immersed. After stirring, the time taken for the resultant emulsion to separate is recorded every 5 minutes until the emulsion separates completely or reduces to 3 ml or less.

14 Done Done Done Done 01:23 01:23 01:24 01:25 54.0 °C 82.0 72% 72% 72% 71% A 54.0 °C 00:00 00:00 00:00 00:00 Abort i You can run multiple tests at the same time 22/Jun/20

Independent control allowing for simultaneous or individual testing








User adjustable test parameters

For more information please visit: www.stanhope-seta.co.uk

Herschel Emulsifier 96700-2


Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412



Ease Of Use

- Fully automated paddle movement with motorised raising and lowering
- Easy sample handling without removing the stirrer paddle
- Self-centering collet to ensure concentricity within 1mm
- Positioning within 6mm from the bottom of the cylinder automatically achieved
- Large LCD touch screen display
- Integrated timer with audible and visual reminders for recording intervals



Enhanced Test Throughput

- 4 independent test stations reduce waiting time
- Rapid bath heat up time

Operator Safety

- Obstruction override if paddles are placed under any resistant load
- Duplex protective viewing glass
- Automatic high temperature cut out
- Overflow prevention
- Automatic low-level liquid trip
- Abort button for emergency all stop



Precision and Accuracy

- Non-reflective enhanced LED lighting for improved measurement precision
- User adjustable sample temperature stabilisation time
- Post stirring, intermediate scraping position ensures accurate readings
- Guaranteed paddle rotation speed, regardless of viscosity
- Large viewing window



For more information please visit: www.stanhope-seta.co.uk

Herschel Emulsifier 96700-2

Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412

Technical Specifications

Herschel Emulsifier				
Bath volume	5 litres			
Bath liquid	Water or White oil			
Sample size	40 ml oil			
	40 ml distilled water			
	1% sodium chloride solution or synthetic seawater			
Test temperatures	54 °C and 82 °C			
Bath temperature stability	±1°C			
Stirrer speed	1500 <u>+</u> 15 rpm			
Voltage options	110 Vac, 50/60 Hz			
	240 Vac, 50/60 Hz			
Current	10 A (max)* - 110 Vac model			
	5 A (max)* - 220 Vac model			
Display	LCD touchscreen			
Size: (H x W x D)	890 x 450 x 450 mm			
Weight (empty)	49.5 kg			

Minimal Servicing

Herschel Emulsifier		
Weekly	Tighten paddle collets	
Annually	Temperature calibration	
	Paddle speed calibration	

The instrument is supplied with the following:

- 4 graduated glass cylinders
- Screen protector
- Spatula

* In-rush current will be higher

Accessories

Part Number		Description		
51000-0		Digital thermometer 6 x 210 mm probe. 5 point UKAS calibration at -50, 0, 50, 100 and 150 °C		
<u>94630-0</u>		Bath oil, blended with anti-oxidant Suitable for temperatures between 40 and 85 °C. Supplied in a 5 litre container		
<u>99960-2</u>		Tachometer Accuracy ±1.75 rpm at 1500 rpm		
<u>11227-0</u>		Synthetic seawater solution Supplied with Certificate of Conformance to ASTM D665		
Agence Nord:Agence Est:ZA Object'lfs Sud - Lot A3Parc Club des Tanneries6 Allée Emilie du Châtelet2 Rue de la Faisanderie14123 lfs67380 Lingolsheimtél : 02.31.34.50.74tél : 03.88.04.01.81fax : 02.31.34.55.17fax : 03.68.93.01.52				
		Sud: /enango. 392 Rue Jean Dausset ARC - BP11575	www.deltalabo.fr info@deltalabo.fr	

84916 Avignon Cédex 9 tél : 04.90.27.17.95 fax : 04.90.27.17.52